



Major Tech (PTY) LTD

TEST REPORT

Page 1 of 1

MODEL:MAJOR MT688 Non-contact Body InfraRed Thermometers

Serial Number:200411566

Report Number:202023252

Calibration Unit Detail (standard)

Manufacturer: HART (InfraRed Temperature Calibration Equipment)

Model Number: 9133

Serial Number: A35050

Environment Temperature:23±2 °C

Humidity: 55±20%RH

Calibration Date:2020-06-03

Issue Date:2020-06-03

Calibrated By:

Issue Name: Peng Xingen

Surface °C

RANGE	INPUT		TOLERANCE			Resolution	READING			RESULTS	STATUS
			(%+dpts)				min	to	max		
0to60°C	36.0	°C	0	+	2	0.1	35.8	to	36.2	36.1	Passed
	39.0	°C	0	+	2	0.1	38.8	to	39.2	39.1	Passed

Body °C

32to42.5°C	37.0	°C	0	+	2	0.1	36.8	to	37.2	37.0	Passed
	39.3	°C	0	+	3	0.1	39.0	to	39.6	39.4	Passed

Conclusion: Pass at Calibration Item.